



NIELSEN-KELLERMAN

21 Creek Circle • Boothwyn, PA 19061 USA
Phone 610.447.1555 • Fax 610.447.1577
info@nkhome.com • www.nkhome.com

DECLARATION OF CONFORMITY

**EC DECLARATION OF CONFORMITY
ACCORDING TO COUNCIL DIRECTIVE 89/336/EEC**

Nielsen-Kellerman Company declares that the following products, manufactured by the Nielsen-Kellerman Company at 21 Creek Circle, Boothwyn, PA 19061 USA:

Product Model	Product Description
Kestrel 4500	Pocket Weather Tracker <i>Combined Wind Speed, Temperature, Relative Humidity and Pressure Measuring Instrument (Anemometer / Thermometer / Hygrometer / Barometer/ Digital Compass)</i>
Kestrel 4500NV	Pocket Weather Tracker (Night Vision Model) <i>Combined Wind Speed, Temperature, Relative Humidity and Pressure Measuring Instrument (Anemometer / Thermometer / Hygrometer / Barometer/ Digital Compass)</i>

Are in conformity with the following standards or normative documents:

EMISSIONS:

EN 61000-6-3:2001 and EN 61326:1997/A1/A2/A3:2003; Radiated disturbance CISPR 22:1997, Class B

IMMUNITY:

EN 61000-6-1:2001 and EN61326: 1997/A1/A2/A3:2003: Electrostatic discharge - IEC 61000-4-2:1995
Radiated RF interference, AM - IEC 61000-4-3:1996
Power frequency magnetic field - IEC 61000-4-8:1994

Name of Authorized Person: Paul Nielsen, Vice-President, Engineering

Signature of Authorized Person:

Test Period: 1-4th August 2006

Certificate Date: 22nd August 2006